



# Certificate of Compliance

**Certificate:** 80081699

**Master Contract:** 301727

**Project:** 80081699

**Date Issued:** 2021-07-30

**Issued To:** Precision ADM  
1595 Buffalo Place Unit A  
Winnipeg, Manitoba, R3T 1L9  
Canada

**Attention:** Matt Wallis

*The products listed below are eligible to bear the CSA Mark shown*

**Issued by:** Wendy Zhu  
Wendy Zhu, Ph. D., P. Eng.



## **PRODUCTS**

CLASS - C720402 - OCCUPATIONAL HEALTH & SAFETY Surgical Non-Powered Air-Purifying Respirators

Surgical Respirator Model Precision Air (Size Small-Medium, and Size Medium-Large) with replaceable Filter Model Viraloc, 95PFE-L3

## **APPLICABLE REQUIREMENTS**

- The Code of Federal Regulation 42 CFR Part 84, Approval of Respiratory Protective Devices, Subpart K (Nonpowered Particulate Respirators).
- The Code of Federal Regulations, Title 16- Commercial Practices, Part 1610, Standard for the Flammability of Clothing Textiles
- ASTM F1862-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood
- CSA TechNote TN202, Filtering Facepiece Respirator Head Strap / Harness Assembly Strength Requirements and Test Procedure



## *Supplement to Certificate of Compliance*

**Certificate:** 80081699

**Master Contract:** 301727

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80081699	2021-07-30	Original certification of Surgical Respirator Model Precision Air (Size Small-Medium, and Size Medium-Large) with replaceable Viraloc Filters, 95PFE-L3 The Code of Federal Regulation 42 CFR Part 84, Subpart K The Code of Federal Regulations, Title 16- Commercial Practices ASTM F1862-17 CSA TechNote TN202 Initial Factory Evaluation (IFE) were conducted under projects 80081700, 80085873 and 80087952. All supporting documents are under project 80081700.